

JPL, 00HW019
Data Validation Reports
LDC# 9432

Wet Chemistry

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 15, 2002
LDC Report Date: December 2, 2002
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5485

Sample Identification

ER-12
MW-12-1
MW-12-2
MW-12-3
MW-12-4
MW-12-5
MW-12-1MS
MW-12-1MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 7196A for Hexavalent Chromium and EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met with the following exceptions:

| Sample | Analyte | Finding | Criteria | Flag | A or P |
|--|-------------|--|---|------|--------|
| ER-12 MW-12-1 MW-12-2 MW-12-3 MW-12-4 MW-12-5 | Perchlorate | Initial calibration was not performed at the required frequency. | Initial calibration must be performed every 6 months. | None | P |

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

Sample ER-12 was identified as an equipment rinsate. No contaminant concentrations were found in this blank.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Wet Chemistry - Data Qualification Summary - SDG 02-5485

| SDG | Sample | Analyte | Flag | A or P | Reason |
|---------|--|-------------|------|--------|---------------------|
| 02-5485 | ER-12 MW-12-1 MW-12-2 MW-12-3 MW-12-4 MW-12-5 | Perchlorate | None | P | Initial calibration |

JPL, 00HW019

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 02-5485

No Sample Data Qualified in this SDG

JPL, 00HW019

Wet Chemistry - Field Blank Data Qualification Summary - SDG 02-5485

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 18, 2002
LDC Report Date: December 2, 2002
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5558

Sample Identification

ER-21
MW-21-1
MW-21-2
MW-21-3
MW-21-4
MW-21-5

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

Sample ER-21 was identified as an equipment rinsate. No perchlorate contaminants were found in this blank.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Perchlorate - Data Qualification Summary - SDG 02-5558

No Sample Data Qualified in this SDG

JPL, 00HW019

Perchlorate - Laboratory Blank Data Qualification Summary - SDG 02-5558

No Sample Data Qualified in this SDG

JPL, 00HW019

Perchlorate - Field Blank Data Qualification Summary - SDG 02-5558

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 21, 2002
LDC Report Date: December 2, 2002
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5592

Sample Identification

ER-19
MW-19-1
MW-19-2
MW-19-3
MW-19-4
MW-19-5
MW-19-5D
MW-19-4MS
MW-19-4MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No perchlorate contaminants were found in the method blanks.

Sample ER-19 was identified as an equipment rinsate. No perchlorate contaminants were found in this blank.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples MW-19-5 and MW-19-5D were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

JPL, 00HW019

Wet Chemistry - Data Qualification Summary - SDG 02-5592

No Sample Data Qualified in this SDG

JPL, 00HW019


Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 02-5592

No Sample Data Qualified in this SDG

JPL, 00HW019

Wet Chemistry - Field Blank Data Qualification Summary - SDG 02-5592

No Sample Data Qualified in this SDG



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

SOTA Environmental
16835 W. Bernardo, Drive, Suite 212
San Diego, CA 92127-1813
ATTN: Ms. Yu Zeng

December 23, 2002

SUBJECT: JPL, 00HW019, Data Validation

Dear Ms. Zeng,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on December 11, 2002. Attachment 1 is a summary of the samples that were reviewed for each analysis.

LDC Project # 9504:

| <u>SDG #</u> | <u>Fraction</u> |
|---|--------------------------|
| 02-5613, 02-5667, 02-5689, 02-5691, 02-5733, 02-5742, 02-5779, 02-5826 | Volatiles, Wet Chemistry |

The data validation was performed under EPA Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996

Please feel free to contact us if you have any questions.

Sincerely,

Stacey A. Mavrakos
Operations Manager/Senior Chemist

LDC #9504 (Sota Environmental-San Diego / JPL, 00HW019)

[illegible]

Shaded cells indicate Level IV validation (all other cells are Level III validation)

JPL, 00HW019
Data Validation Reports
LDC# 9504

Volatiles

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 22, 2002
LDC Report Date: December 17, 2002
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5613

Sample Identification

ER-23
MW-23-1
MW-23-2
MW-23-3
MW-23-3D
TB-23

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-23 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Sampling Date | Compound | Concentration | Associated Samples |
|---------------|---------------|--------------------|---------------|--|
| TB-23 | 10/22/02 | Methylene chloride | 1 ug/L | ER-23 MW-23-1 MW-23-2 MW-23-3 MW-23-3D |

Sample ER-23 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples MW-23-3 and MW-23-3D were identified as field duplicates. No volatiles were detected in any of the samples.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 02-5613

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5613

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 02-5613

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 23, 2002
LDC Report Date: December 17, 2002
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5667

Sample Identification

ER-22
FB-22
MW-22-1
MW-22-2
MW-22-3
TB-22

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

| Date | Compound | %D | Associated Samples | Flag | A or P |
|----------|---------------------|-------|----------------------------|---|--------|
| 10/31/02 | 2,2-Dichloropropane | 38.80 | All samples in SDG 02-5667 | J (all detects) UJ (all non-detects) | A |

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-22 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Sampling Date | Compound | Concentration | Associated Samples |
|---------------|---------------|--------------------|---------------|---|
| TB-22 | 10/23/02 | Methylene chloride | 1.0 ug/L | ER-22 FB-22 MW-22-1 MW-22-2 MW-22-3 |

Sample ER-22 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

Sample FB-22 was identified as a field blank. No volatile contaminants were found in this blank.

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019**Volatiles - Data Qualification Summary - SDG 02-5667**

| SDG | Sample | Compound | Flag | A or P | Reason |
|---------|--|---------------------|---|--------|--------------------------------|
| 02-5667 | ER-22 FB-22 MW-22-1 MW-22-2 MW-22-3 TB-22 | 2,2-Dichloropropane | J (all detects) UJ (all non-detects) | A | Continuing calibration (%D) |

JPL, 00HW019**Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5667**

No Sample Data Qualified in this SDG

JPL, 00HW019**Volatiles - Field Blank Data Qualification Summary - SDG 02-5667**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 24, 2002
LDC Report Date: December 17, 2002
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5689

Sample Identification

MW-6
MW-13
TB-6

Introduction

This data review covers 3 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

| Date | Compound | %D | Associated Samples | Flag | A or P |
|----------|---------------------|-------|-----------------------------|---|--------|
| 10/31/02 | 2,2-Dichloropropane | 38.80 | MW-6 TB-6 02G4474MB01 | J (all detects) UJ (all non-detects) | A |

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-6 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Sampling Date | Compound | Concentration | Associated Samples |
|----------------------|----------------------|--------------------|----------------------|---------------------------|
| TB-6 | 10/24/02 | Methylene chloride | 3.0 ug/L | MW-6 MW-13 |

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 02-5689

| SDG | Sample | Compound | Flag | A or P | Reason |
|---------|--------------|---------------------|---|--------|--------------------------------|
| 02-5689 | MW-6 TB-6 | 2,2-Dichloropropane | J (all detects) UJ (all non-detects) | A | Continuing calibration (%D) |

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5689

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 02-5689

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 25, 2002
LDC Report Date: December 17, 2002
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5691

Sample Identification

MW-5
TB-5

Introduction

This data review covers 2 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

| Date | Compound | %D | Associated Samples | Flag | A or P |
|----------|---------------------|-------|----------------------------|---|--------|
| 10/31/02 | 2,2-Dichloropropane | 38.80 | All samples in SDG 02-5691 | J (all detects) UJ (all non-detects) | A |

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-5 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Sampling Date | Compound | Concentration | Associated Samples |
|----------------------|----------------------|--------------------|----------------------|---------------------------|
| TB-5 | 10/25/02 | Methylene chloride | 2.2 ug/L | MW-5 |

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks with the following exceptions:

| Sample | Compound | Reported Concentration | Modified Final Concentration |
|---------------|--------------------|-------------------------------|-------------------------------------|
| MW-5 | Methylene chloride | 1.3 ug/L | 1.3U ug/L |

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 02-5691

| SDG | Sample | Compound | Flag | A or P | Reason |
|---------|--------------|---------------------|---|--------|--------------------------------|
| 02-5691 | MW-5 TB-5 | 2,2-Dichloropropane | J (all detects) UJ (all non-detects) | A | Continuing calibration (%D) |

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5691

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 02-5691

| SDG | Sample | Compound | Modified Final Concentration | A or P |
|---------|--------|--------------------|---------------------------------|--------|
| 02-5691 | MW-5 | Methylene chloride | 1.3U ug/L | A |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 28, 2002
LDC Report Date: December 17, 2002
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5733

Sample Identification

ER-3
MW-3-2
MW-3-3
MW-3-4
MW-3-4D
TB-3
MW-3-3MS
MW-3-3MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

| Date | Compound | %D | Associated Samples | Flag | A or P |
|----------|---------------------|-------|----------------------------|---|--------|
| 10/31/02 | 2,2-Dichloropropane | 38.80 | All samples in SDG 02-5733 | J (all detects) UJ (all non-detects) | A |

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-3 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample ER-3 was identified as an equipment rinsate. No volatile contaminants were found in this blank with the following exceptions:

| Equipment Rinsate ID | Sampling Date | Compound | Concentration | Associated Samples |
|----------------------|---------------|--------------------|---------------|---------------------------------------|
| ER-3 | 10/28/02 | Methylene chloride | 0.7 ug/L | MW-3-2 MW-3-3 MW-3-4 MW-3-4D |

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater ($>10X$ for common contaminants, $>5X$ for other contaminants) than the concentrations found in the associated field blanks with the following exceptions:

| Sample | Compound | Reported Concentration | Modified Final Concentration |
|---------|--------------------|------------------------|------------------------------|
| MW-3-2 | Methylene chloride | 0.3 ug/L | 1U ug/L |
| MW-3-3 | Methylene chloride | 0.3 ug/L | 1U ug/L |
| MW-3-4 | Methylene chloride | 0.5 ug/L | 1U ug/L |
| MW-3-4D | Methylene chloride | 0.7 ug/L | 1U ug/L |

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples MW-3-4 and MW-3-4D were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

| Compound | Concentration (ug/L) | | RPD |
|-------------------------|----------------------|---------|-----|
| | MW-3-4 | MW-3-4D | |
| Methylene chloride | 0.5 | 0.7 | 33 |
| Methyl-tert-butyl ether | 5.5 | 1U | 200 |
| Toluene | 0.5 | 0.5U | 200 |

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 02-5733

| SDG | Sample | Compound | Flag | A or P | Reason |
|---------|---|---------------------|---|--------|-----------------------------|
| 02-5733 | ER-3 MW-3-2 MW-3-3 MW-3-4 MW-3-4D TB-3 | 2,2-Dichloropropane | J (all detects) UJ (all non-detects) | A | Continuing calibration (%D) |

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5733

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 02-5733

| SDG | Sample | Compound | Modified Final Concentration | A or P |
|---------|---------|--------------------|------------------------------|--------|
| 02-5733 | MW-3-2 | Methylene chloride | 1U ug/L | A |
| 02-5733 | MW-3-3 | Methylene chloride | 1U ug/L | A |
| 02-5733 | MW-3-4 | Methylene chloride | 1U ug/L | A |
| 02-5733 | MW-3-4D | Methylene chloride | 1U ug/L | A |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 29, 2002
LDC Report Date: December 17, 2002
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5742

Sample Identification

ER-24
MW-24-1
MW-24-2
MW-24-3
TB-24

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-24 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample ER-24 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 02-5742

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5742

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 02-5742

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 30, 2002
LDC Report Date: December 17, 2002
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 02-5779

Sample Identification

ER-4
MW-4-1
MW-4-2
MW-4-3
MW-4-2D
TB-4
MW-4-3MS
MW-4-3MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-4 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Sampling Date | Compound | Concentration | Associated Samples |
|---------------|---------------|--------------------|---------------|---|
| TB-4 | 10/30/02 | Methylene chloride | 1.1 ug/L | ER-4 MW-4-1 MW-4-2 MW-4-3 MW-4-2D |

Sample ER-4 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater ($>10X$ for common contaminants, $>5X$ for other contaminants) than the concentrations found in the associated field blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples MW-4-2 and MW-4-2D were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

| Compound | Concentration (ug/L) | | RPD |
|--------------------|----------------------|---------|-----|
| | MW-4-2 | MW-4-2D | |
| Chloroform | 0.4 | 0.4 | 0 |
| 1,1-Dichloroethane | 0.5 | 0.6 | 18 |
| Tetrachloroethene | 0.5 | 0.8 | 46 |
| Trichloroethene | 1.1 | 1.7 | 43 |

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 02-5779

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5779

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 02-5779

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: October 31, 2002
LDC Report Date: December 17, 2002
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 02-5826

Sample Identification

ER-14
MW-14-1
MW-14-2
MW-14-3
MW-14-4
MW-14-5
TB-14

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-14 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample ER-14 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 02-5826

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5826

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 02-5826

No Sample Data Qualified in this SDG